

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Cher Snyder, Assistant DEO
Engineering & Compliance

LABORATORY NO 1606727

DATE RECEIVED 03/08/2016

SAMPLE DESCRIBED AS

24 hour sample
canister # 54549

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION

Porter Ranch/Castlebay
School
Porter Ranch, CA 91326

ST NO / PROJECT NA

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54549</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	396
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 3/10/16

Approved By: 

Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1606727</u>
24 hour sample canister # 54549	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	9.78	-6.00	PASS
CH ₄ , ppmvC	0.3	10.17	10.40	2.29	PASS
CO ₂ , ppmvC	0.4	10.38	10.46	0.75	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.36	-5.83	PASS
NM/NEOC, ppmvC	0.2	10.64	10.16	-4.50	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	9637	-4.59	PASS
CH ₄ , ppmvC	0.3	9950	9988	0.38	PASS
CO ₂ , ppmvC	0.4	10100	9629	-4.66	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	9817	-1.24	PASS
NM/NEOC, ppmvC	0.2	10000	9726	-2.74	PASS

DATE ANALYZED 03/08/2016
REFERENCE NO: 16QM2AA
 QM2-101-85

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 INV
 LAF
 LABO

WO #: 1606727



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 3/8/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 3/8/16 Time: 9:25am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem-A	54549	3/7/16 / 00:00 / 24 hours	-30"	+10.5
Porter Ranch / Castlebay Elem-B	54676	3/7/16 / 00:00 / 24 hours	-30"	>+15

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhongqian</i>	<i>JL</i>	SCAQMD Lab	3/8/16	11:49

Remarks: 1:3 scheduled samples from trailer at Castlebay
Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA
A: AQMD 0013772; B: AQMD 0016503