

1778

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Cher Snyder, Assistant DEO Engineering & Compliance	LABORATORY NO	<u>1607611</u>
		DATE RECEIVED	<u>03/17/2016</u>
SAMPLE DESCRIBED AS	One Summa Canister, 24 h sample	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
SAMPLING LOCATION	Highlands Community 12378 High Glen Way Northridge, CA 91326	ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54570</u>
Pressure (Torr)	509
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	394
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 3/18/16

Approved By: 

Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1607611</u>
One Summa Canister, 24 h sample	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	9.85	-5.27	PASS
CH ₄ , ppmvC	0.3	10.17	10.65	4.70	PASS
CO ₂ , ppmvC	0.4	10.38	10.46	0.81	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.16	-7.67	PASS
NM/NEOC, ppmvC	0.2	10.64	10.23	-3.87	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	9654	-4.42	PASS
CH ₄ , ppmvC	0.3	9950	10007	0.57	PASS
CO ₂ , ppmvC	0.4	10100	9644	-4.51	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	9825	-1.15	PASS
NM/NEOC, ppmvC	0.2	10000	9736	-2.64	PASS

DATE ANALYZED	03/17/2016
REFERENCE NO:	16QM2AA QM2-101-89

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AL
LABORATORY

WO #: 1607611



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 3/17/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 3/17/16 Time: 8:50am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54570	3-16-16 / 00:00 / 24 hours	-30"	-11"

Relinquished by	Received by	Firm/Agency	Date	Time
zhang qian	<i>[Signature]</i>	SCAQMD Lab	3/17/16	12:20pm

Remarks Daily scheduled samples from trailer

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)

Pi : 509.1

Pf : 850.2

MR 3/17/2016