

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Cher Snyder, Assistant DEO
Engineering & Compliance

LABORATORY NO 1608812

DATE RECEIVED 03/29/2016

SAMPLE DESCRIBED AS

24 hour sample
canister # E4271

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION


Porter Ranch
Elementary School

ST NO / PROJECT NA

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>E4271</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	412
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/6/16

Approved By: 

Solomon Teffera, Acting Manager
Laboratory Services
909-396-2199

Handwritten: 144
3/30/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1608812</u>
24 hour sample canister # E4271	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.09	8.78	PASS
CH ₄ , ppmvC	0.3	2.02	2.10	3.92	PASS
CO ₂ , ppmvC	0.3	1.57	1.94	23.58	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.17	7.03	PASS
NM/NEOC, ppmvC	0.4	2.03	2.12	4.67	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1054	1.76	PASS
CH ₄ , ppmvC	0.3	1068	1049	-1.78	PASS
CO ₂ , ppmvC	0.3	1022	1041	1.81	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1042	-0.17	PASS
NM/NEOC, ppmvC	0.4	1024	1017	-0.67	PASS

DATE ANALYZED 03/29/2016
REFERENCE NO: 16QM2AB
 QM2-101-94

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 IN
 LA
LABC

WO #: 1608812



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 3/29/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 3/29/16 Time: 09:35

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)	E4271	3/28/16 / 00:00 / 24 hours	<-30"	+1

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>JL</i>	SCAQMD Lab	3/29/16	12:03

Remarks: 1:3 scheduled samples from trailer at PRCS
Porter Ranch Community School Elementary – 12450 Mason Ave, Porter Ranch, CA 91326
GPS (34.293369, -118.580505)