

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

| | | | |
|----------------------------|---|------------------------|----------------------|
| TO | Jason Low Atmospheric Measurements Manager Science and Technology Advancement | LABORATORY NO | <u>1613013</u> |
| SAMPLE DESCRIBED AS | 24 hour sample Canister 54715 | DATE RECEIVED | <u>05/10/2016</u> |
| SAMPLING LOCATION | Reseda Station 18328 Gault St Los Angeles, CA 91335 | FACILITY ID NO | <u>NA</u> |
| | | REQUESTED BY | <u>Sumner Wilson</u> |
| | | ST NO / PROJECT | <u>NA</u> |

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

| | |
|-----------------------|--------------|
| Type | Canister |
| Number | <u>54715</u> |
| Pressure (Torr) | 1388 |
| CO, ppm | < 3 |
| CH ₄ , ppm | 2 |
| CO ₂ , ppm | 436 |
| Ethane, ppmvC | < 1 |
| NM/NEOC, ppmvC | < 1 |

Date Approved: 5/12/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

945-1116

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

| | | |
|----------------------------------|----------------------|----------------------|
| SAMPLE DESCRIBED AS | LABORATORY NO | <u>1613013</u> |
| 24 hour sample Canister 54715 | REQUESTED BY | <u>Sumner Wilson</u> |

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

| | | Theoretical | Measured | Percent Difference | QC Limit ±5% or ± 1 |
|---------------------------------------|-----|-------------|----------|--------------------|------------------------|
| CC106783 | MDL | | | | |
| CO, ppmvC | 0.3 | 1.92 | 2.00 | 4.20 | PASS |
| CH ₄ , ppmvC | 0.3 | 2.02 | 1.98 | -2.02 | PASS |
| CO ₂ , ppmvC | 0.3 | 1.57 | 1.93 | 23.21 | PASS |
| C ₂ H ₄ , ppmvC | 0.4 | NA | NA | NA | NA |
| C ₂ H ₆ , ppmvC | 0.4 | 2.03 | 1.92 | -5.43 | PASS |
| NM/NEOC, ppmvC | 0.4 | 2.03 | 2.00 | -1.52 | PASS |

| | | Theoretical | Measured | Percent Difference | QC Limit ±5% or ± 1 |
|---------------------------------------|-----|-------------|----------|--------------------|------------------------|
| CC12628 | MDL | | | | |
| CO, ppmvC | 0.3 | 1036 | 1071 | 3.42 | PASS |
| CH ₄ , ppmvC | 0.3 | 1068 | 1065 | -0.28 | PASS |
| CO ₂ , ppmvC | 0.3 | 1022 | 1056 | 3.36 | PASS |
| C ₂ H ₄ , ppmvC | 0.4 | NA | NA | NA | NA |
| C ₂ H ₆ , ppmvC | 0.4 | 1044 | 1058 | 1.31 | PASS |
| NM/NEOC, ppmvC | 0.4 | 1024 | 1035 | 1.13 | PASS |

DATE ANALYZED 05/10/2016
REFERENCE NO: 16QM2AB
 QM2-101-102

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIST
 INVC
 LAP
 LABOR

WO #: 1613013



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: _____ Southern California Gas Co. _____ I.D. No. _____
 Source Address: _____ 12801 Tampa Ave _____ City: _____ Porter Ranch _____
 Mailing Address: _____ City: _____ Zip: _____ 91326 _____
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: _____ Sumner Wilson _____ Date: _____ 5/10/16 _____
 Approved by: _____ Jason Low _____ Office: _____ Budget #: _____ 44716 _____
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: _____ Qian Zhou _____ Date: _____ 5/10/16 _____ Time: _____ 09:50pm _____

REQUESTED ANALYSIS: PAMS analysis

| City/Location | Can# | Start day / time/ duration | Start vac | End Press |
|----------------|-------|----------------------------|-----------|-----------|
| Reseda Station | 54715 | 5/9/16 / 00:00 / 24 hours | <-30" | +9.5 |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

| Relinquished by | Received by | Firm/Agency | Date | Time |
|------------------|-------------|-------------|---------|-------|
| <i>zhongqian</i> | <i>JL</i> | SCAQMD Lab | 5/10/16 | 11:25 |
| | | | | |
| | | | | |

Remarks: 1:3 scheduled samples from station at Reseda
 Reseda Station – 18328 Gault St, Los Angeles, CA 91335
 GPS (34.199225, -118.532743)