

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low ADEO Science and Technology Advancement	LABORATORY NO	<u>1627726</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54730	DATE RECEIVED	<u>10/11/2016</u>
SAMPLING LOCATION	Highlands Community Center	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>Highlands Community</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54730</u>
Pressure (Torr)	760
CO, ppm	4
CH ₄ , ppm	3
CO ₂ , ppm	410
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 10/22/16²¹

Approved By: 

Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1627726</u>
24 hour sample Canister 54730	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.34	21.81	PASS
CH ₄ , ppmvC	0.3	2.02	2.21	9.00	PASS
CO ₂ , ppmvC	0.4	1.57	2.31	47.06	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.97	-2.92	PASS
NM/NEOC, ppmvC	0.5	2.03	2.41	18.77	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1053	1.61	PASS
CH ₄ , ppmvC	0.3	1068	1048	-1.89	PASS
CO ₂ , ppmvC	0.4	1022	1035	1.29	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1036	-0.80	PASS
NM/NEOC, ppmvC	0.5	1024	1008	-1.51	PASS

DATE ANALYZED 10/11/2016
REFERENCE NO: 16QM2AD
 QM2-101-126

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 INV
 LAI
LABO.

WO #: 1627726



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 10/11/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 10/11/16 Time: 12:20pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54730	10/9/16 / 00:00 / 24 hours	-30"	+11

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhongqian</i>	<i>[Signature]</i>	SCAQMD Lab	10/11/2014	15:08

Remarks 1:6 scheduled samples from trailer
 Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)
 Sampler sn 5323